## How to Administer an

## **Elastomeric Infusion Pump**







Administration of an Elastomeric Infusion Pump	
Proce	dure to attach an elastomeric infusion pump to an existing IV catheter
9. Pre	epare your syringes.
a)	Remove the syringe cap.
b)	) Gently tap the syringe to move air to the tip.
c)	Push the plunger to remove air and extra solution. (if difficult to push, pull back slightly to break the seal)
d	) Replace the syringe cap until it is to be used.
10. Flu	sh the line.
a)	Remove the existing alcohol disinfectant cap and discard.
b)	) Attach the syringe firmly to the IV needleless connector using a push and twist motion
c)	Open the clamp.
d	) Inject the solution.
e)	Close the Clamp.
f)	Remove and discard the syringe.
g)	Attach a new alcohol disinfectant cap after the final flush.

<section-header><section-header><section-header><section-header><section-header><section-header><section-header><section-header><section-header><section-header><list-item><list-item><list-item><section-header>

## Administration of an Elastomeric Infusion Pump

## Documentation

Documentation in the medical records includes, but is not limited to:

- Date and time
- Solution
- Rate and method of infusion
- Complications and interventions
- Patient response to procedure and/or solution
- Patient/significant other education
- Communication with prescribing physician/LIP
- Communication with supervisor

7

