

Programming a New Cyclic Infusion



1. Select "Standard Library"



2. Type "TP" When Prompted



3. Select "TPN"



4. Select "Continuous"



5. Enter Volume and Time



6. Review Ramp Time